K112602

NOV - 7 2011



# Tab 21 PREMARKET NOTIFICATION 510(K) SUMMARY

This summary of 510(k) safety and effectiveness information is being submitted in accordance with the requirements of SMDA 1990 and 21 CFR §807.92.

The assigned 510(k) number is:

**Manufacturer:**

SonoScape Company Limited

Address: 4/F., Yizhe Building, Yuquan Road,Nanshan, Shenzhen 518051, P.R.China

Tel: (86) 755-26722890

Fax: (86) 755-26722850

**Contact Person:** Chen Zhiqiang

Date Prepared: August 31, 2011

**Name of the device:**

*** Trade/Proprietary Name:**

S6 Diagnostic Ultrasound System

*** Common Name:** Diagnostic Ultrasound System and Transducers

*** Classification:**

Regulatory Class: II

Review Category: Tier II

21 CFR 892.1550 Ultrasonic Pulsed Doppler Imaging System (90-IYN)

21 CFR 892.1560 Ultrasonic Pulsed Echo Imaging System (90-IYO)

21 CFR 892.1570 Diagnostic Ultrasound Transducer (90-ITX)

**Legally Marketed Predicate Device:**

S20 Digital Doppler Ultrasound System and Transducers –K110510

**Device Description:**

The SonoScape S6 Diagnostic Ultrasound System is an integrated preprogrammed

color ultrasound imaging system, capable of producing high detail resolution intended for clinical diagnostic imaging applications.

The all digital architecture with progressive dynamic receive focusing allows the system to maximize the utility of all imaging transducers to enhance the diagnostic utility and confidence provided by the system. The exam dependent default setting allows the user to have minimum adjustment for imaging the patient, while the in-depth soft-menu control allows the advanced user to set the system for different situations. The architecture allows cost-effective system integration to a variety of upgrade-able options and features.

This SonoScape system is a general purpose, software controlled, diagnostic ultrasound system. Its basic function is to acquire ultrasound data and display the image in B-Mode (including Tissue Harmonic Image), M-Mode, TDI, Color-Flow Doppler, Pulsed Doppler and Power Doppler, or a combination of these modes, 3D/4D.

**Intended Use:**

The device is a general-purpose ultrasonic imaging instrument intended for use by a qualified physician for evaluation of Fetal, Abdominal, Pediatric, Small Organ (breast, testes, thyroid), Cephalic(neonatal and adult), Trans-rectal, Trans-vaginal, Peripheral Vascular, Musculo-skeletal (Conventional and Superficial), Cardiac (neonatal and adult), OB/Gyn and Urology.

## Probe Information:

**Tab 21.1 Probe information**

| No. | Probe | Type | Frequency Range | Intended Use |
|---|---|---|---|---|
| 1 | 2P1 | Phased Array | 2.0-4.0 MHz | Abdominal<br>Neonatal Cephalic<br>Adult Cephalic<br>Cardiac Adult<br>Cardiac Pediatric |
| 2 | 5P1 | Phased Array | 3.0-7.0 MHz | Pediatric<br>Neonatal Cephalic<br>Cardiac Pediatric |
| 3 | 6V1 | Micro-curved Array | 4.0-8.0 MHz | Trans-rectal<br>Trans-vaginal |
| 4 | 6V3 | Micro-curved Array | 5.0-9.0 MHz | Trans-rectal<br>Trans-vaginal |
| 5 | EC9-5 | Micro-curved Array | 5.0-9.0 MHz | Trans-rectal<br>Trans-vaginal |
| 6 | C611 | Micro-curved Array | 4.0-8.0 MHz | Abdominal<br>Pediatric<br>Neonatal Cephalic<br>Cardiac Pediatric |
| 7 | C344 | Curved Array | 2.0-5.0 MHz | Fetal / Abdominal/ Ob/GYN |
| 8 | C362 | Curved Array | 2.0-6.0 MHz | Fetal / Abdominal/ Ob/GYN |
| 9 | VC6-2 | Curved Array | 2.0-6.0 MHz | Fetal / Abdominal/ Ob/GYN |
| 10 | L741 | Linear Array | 5.0-10.0 MHz | Small Organ (reast, thyroid, testes)<br>Musculo-skeletal (Conventional)<br>Peripheral vessel |
| 11 | L742 | Linear Array | 5.0-12.0 MHz | Small Organ (reast, thyroid, testes)<br>Musculo-skeletal (Conventional)<br>Musculo-skeletal (Superficial)<br>Peripheral vessel |
| 12 | L743 | Linear Array | 5.0-10.0 MHz | Small Organ (reast, thyroid, testes)<br>Musculo-skeletal (Conventional)<br>Musculo-skeletal (Superficial)<br>Peripheral vessel |

## Safety Considerations:

The S6 Diagnostic Ultrasound System with added transducer incorporates the same fundamental technology as the predicate device. The device has been tested as Track 3 Device per the FDA Guidance document "Information for Manufacturers Seeking Marketing Clearance of Diagnostic Ultrasound Systems and Transducers" issued September 9, 2008. The acoustic output is measured and calculated per

NEMA UD 2: 2004 Acoustic Output Measurement Standard for Diagnostic Ultrasound Equipment and NEMA UD3: 2004 Standards for Real-time Display of Thermal and Mechanical Acoustic Output Indices on Diagnostic Ultrasound Equipment. The device conforms to applicable medical device safety standards, such as IEC 60601-1, IEC 60601-1-1, IEC 60601-1-2, IEC 60601-2-37, ISO 10993-5 and ISO 10993-10.

**Testing:**

Laboratory testing was conducted to verify that the S6 Diagnostic Ultrasound System with added transducer met all design specification and was substantially equivalent to the currently marketed Predicate Device as above. The device has been found to conform to applicable medical device safety standards in regards to thermal, mechanical and electrical safety as well as biocompatibility. Acoustic output is measured and calculated according to "Acoustic Output Measuring Standard for Diagnostic Ultrasound Equipment".

**Tab 21.2 Applicable Safety Standards**

| Standards No. | Standards Title | Version | Date |
|---|---|---|---|
| IEC 60601-1 | IEC 60601-1, Medical Electrical Equipment - Part 1: General Requirements for Safety, 1988; Amendment 1, 1991-11, Amendment 2, 1995. | 1988 | 10/31/2005 |
| IEC 60601-1-2 | IEC 60601-1-2, (Second Edition, 2001), Medical Electrical Equipment - Part 1-2: General Requirements for Safety - Collateral Standard: Electromagnetic Compatibility -- Requirements and Tests. | 2007 | 07/31/2008 |
| IEC 60601-2-37 | IEC 60601-2-37 (2004) (2005) Amendment 2, Medical electrical equipment - Part 2-37: Particular requirements for the basic safety and essential performance of ultrasonic medical diagnostic and monitoring equipment. | 2007 | 09/08/2009 |
| NEMA UD 2 | NEMA UD 2-2004, Acoustic Output Measurement Standard for Diagnostic Ultrasound Equipment Version 3. | 2004 | 09/08/2009 |
| ISO 10993-5 | ISO 10993-5:1999, Biological evaluation of medical devices -- Part 5: Tests for In Vitro cytotoxicity. | 2009 | 09/12/2007 |
| 10993-10 | ISO 10993-10:2002, Biological evaluation of medical devices - Part 10: Tests for irritation and delayed-type hypersensitivity. | 2002 | 09/12/2007 |

**Conclusion:**

The conclusions drawn from testing of the S6 Diagnostic Ultrasound System with transducer demonstrate that the device is as safe and effective as the legally marketed predicate devices.

DEPARTMENT OF HEALTH & HUMAN SERVICES

Public Health Service

Food and Drug Administration
10903 New Hampshire Avenue
Silver Spring, MD 20993

NOV - 7 2011

SonoScape Company Limited
% Ms. Diana Hong
General Manager
Mid-Link Consulting Co., Ltd.
P.O. Box 237-023
Shanghai, 200237
CHINA

Re: K112602
Trade/Device Name: S6 Diagnostic Ultrasound System
Regulation Number: 21 CFR 892.1550
Regulation Name: Ultrasonic pulsed doppler imaging system
Regulatory Class: II
Product Code: IYN, IYO, ITX
Dated: September 6, 2011
Received: September 7, 2011

Dear Ms. Hong:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and we have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

This determination of substantial equivalence applies to the following transducers intended for use with the S6 Diagnostic Ultrasound System, as described in your premarket notification:

Transducer Model Number

2P1 Phased Array
5P1 Phased Array
6V1 Micro-curved Array
6V3 Micro-curved Array
EC9-5 Micro-curved Array
C611 Micro-curved Array
C362 Curved Array
C344 Curved Array

VC6-2 Curved Array
L743 Linear Array
L741 Linear Array
L742 Linear Array

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 895. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

This letter will allow you to begin marketing your device as described in your premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus permits your device to proceed to market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please go to http://www.fda.gov/AboutFDA/CentersOffices/CDRH/CDRHOffices/ucm115809.htm for the Center for Devices and Radiological Health's (CDRH's) Office of Compliance. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

If you have any questions regarding the content of this letter, please contact Shahram Vaezy at (301) 796-6242.

Sincerely Yours,

Mary S. Pastel, Sc.D.
Director
Division of Radiological Devices
Office of In Vitro Diagnostic Device
Evaluation and Safety
Center for Devices and Radiological Health

Enclosure(s)

## 4.4 Tab 3 Indications For Use

510(k) Number:

Device Name: S6 Diagnostic Ultrasound System

Indications for Use: The SonoScape S6 device is a general-purpose ultrasonic imaging instrument intended for use by a qualified physician for evaluation of Fetal, Abdominal, Pediatric, Small Organ (breast, testes, thyroid), Cephalic(neonatal and adult), Trans-rectal, Trans-vaginal, Peripheral Vascular, Musculo-skeletal (Conventional and Superficial), Cardiac (neonatal and adult), OB/Gyn and Urology.

Prescription Use ___X___ AND/OR Over-The-Counter Use _____
(Part 21 CFR 801 Subpart D) (21 CFR 807 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of In Vitro Diagnostic Devices (OIVD)

Mary S Pastel

(Division Sign-Off)
Division of Radiological Devices
Office of *In Vitro* Diagnostic Device Evaluation and Safety

510(k) Number K112602

## Diagnostic Ultrasound Indications for Use Form

**System:** **SonoScape S6**
**Diagnostic Ultrasound Pulsed Echo System**
**Diagnostic Ultrasound Pulsed Doppler Imaging System**

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

| Clinical Application | | Mode of Operation | | | | | | | |
|---|---|---|---|---|---|---|---|---|---|
| General (TRACK 1 ONLY) | Specific (TRACKS 1 & 3) | B | M | PWD | CWD | Color Doppler | Power (Amplitude) Doppler | Other* Combined | Other* Specify |
| Ophthalmic | Ophthalmic | | | | | | | | |
| Fetal Imaging& Other | Fetal | N | N | N | | N | N | Note 1 | Notes 2,4,5 |
| | Abdominal | N | N | N | | N | N | Note 1 | Notes 2,4,5 |
| | Intra-operative Specify | | | | | | | | |
| | Intra-operative Neuro | | | | | | | | |
| | Laparoscopic | | | | | | | | |
| | Pediatric | N | N | N | | N | N | Note 1 | Notes 2,4 |
| | Small Organ (specify) | N | N | N | | N | N | Note 1 | Notes 2,4,6 |
| | Neonatal Cephalic | N | N | N | N | N | N | Note 1 | Notes 2,3,4 |
| | Adult Cephalic | N | N | N | N | N | N | Note 1 | Notes 2,3,4 |
| | Trans-rectal | N | N | N | | N | N | Note 1 | Notes 2,4 |
| | Trans-vaginal | N | N | N | | N | N | Note 1 | Notes 2,4 |
| | Trans-urethral | | | | | | | | |
| | Trans-esoph.(non-Card) | | | | | | | | |
| | Musculo-skeletal (Conventional) | N | N | N | | N | N | Note 1 | Notes 2,4 |
| | Musculo-skeletal (Superficial) | N | N | N | | N | N | Note 1 | Notes 2,4 |
| | Intravascular | | | | | | | | |
| | Other (Ob/GYN) | N | N | N | | N | N | Note 1 | Notes 2,4,5 |
| Cardiac | Cardiac Adult | N | N | N | N | N | N | Note 1 | Notes 2,3,4 |
| | Cardiac Pediatric | N | N | N | N | N | N | Note 1 | Notes 2,3,4 |
| | Intravascular(Cardiac) | | | | | | | | |
| | Trans-esoph.(Cardiac) | | | | | | | | |
| | Intra-cardiac | | | | | | | | |
| | Other (specify) | | | | | | | | |
| Peripheral Vessel | Peripheral vessel | N | N | N | | N | N | Note 1 | Notes 2,4 |
| | Other (specify) | | | | | | | | |

**N = new indication;** **P = previously cleared by FDA;** **E = added under this appendix**

Note 1: Other Combined includes: B/M; B/PWD; B/THI; M/Color M ; B/Color Doppler; B/Color Doppler/PWD; B/Power Doppler/PWD

Note 2: Tissue Harmonic Imaging. The feature does not use contrast agents

Note 3: TDI Note 4: 3D Note 5: 4D

Note 6: Small Organ: breast, thyroid, testes

Prescription Use ____X___ AND/OR Over-The-Counter Use _____
(Part 21 CFR 801 Subpart D) (21 CFR 807 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of In Vitro Diagnostic Devices (OIVD)

Mary S Patel
(Division Sign-Off)
Division of Radiological Devices
Office of In Vitro Diagnostic Device Evaluation and Safety

510K K112602

## Diagnostic Ultrasound Indications for Use Form

Transducer: 2P1 Phase Array
Diagnostic Ultrasound Transducer

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

| Clinical Application | | Mode of Operation | | | | | | | |
|---|---|---|---|---|---|---|---|---|---|
| General (TRACK 1 ONLY) | Specific (TRACKS 1 & 3) | B | M | PWD | CWD | Color Doppler | Power (Amplitude) Doppler | Other* Combined | Other* Specify |
| Ophthalmic | Ophthalmic | | | | | | | | |
| Fetal Imaging& Other | Fetal | | | | | | | | |
| | Abdominal | N | N | N | | N | N | Note 1 | Notes 2,4 |
| | intra-operative Specify | | | | | | | | |
| | Intra-operative Neuro | | | | | | | | |
| | Laparoscopic | | | | | | | | |
| | Pediatric | | | | | | | | |
| | Small Organ (specify) | | | | | | | | |
| | Neonatal Cephalic | N | N | N | N | N | N | Note 1 | Notes 2,3,4 |
| | Adult Cephalic | N | N | N | N | N | N | Note 1 | Notes 2,3,4 |
| | Trans-rectal | | | | | | | | |
| | Trans-vaginal | | | | | | | | |
| | Trans-urethral | | | | | | | | |
| | Trans-esoph.(non-Card) | | | | | | | | |
| | Musculo-skeletal (Conventional) | | | | | | | | |
| | Musculo-skeletal (Superficial) | | | | | | | | |
| | Intravascular | | | | | | | | |
| | Other (Ob/GYN) | | | | | | | | |
| Cardiac | Cardiac Adult | N | N | N | N | N | N | Note 1 | Notes 2,3,4 |
| | Cardiac Pediatric | N | N | N | N | N | N | Note 1 | Notes 2,3,4 |
| | Intravascular(Cardiac) | | | | | | | | |
| | Trans-esoph.(Cardiac) | | | | | | | | |
| | Intra-cardiac | | | | | | | | |
| | Other (specify) | | | | | | | | |
| Peripheral Vessel | Peripheral vessel | | | | | | | | |
| | Other (specify) | | | | | | | | |

**N = new indication; P = previously cleared by FDA; E = added under this appendix**

Note 1: Other Combined includes: B/M; B/PWD; B/THI; M/Color M ; B/Color Doppler; B/Color Doppler/PWD; B/Power Doppler/PWD

Note 2: Tissue Harmonic Imaging. The feature does not use contrast agents

Note 3: TDI Note 4: 3D Note 5: 4D

Note 6: Small Organ: breast, thyroid, testes

Prescription Use ___X___ AND/OR Over-The-Counter Use _____
(Part 21 CFR 801 Subpart D) (21 CFR 807 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of In Vitro Diagnostic Devices (OIVD)

Indications For Use

Mary S Pastel
(Division Sign-Off)
Division of Radiological Devices
Office of In Vitro Diagnostic Device Evaluation and Safety

510K K112602

## Diagnostic Ultrasound Indications for Use Form

Transducer: 5P1 Phase Array
Diagnostic Ultrasound Transducer

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

| Clinical Application | | Mode of Operation | | | | | | | |
|---|---|---|---|---|---|---|---|---|---|
| General (TRACK 1 ONLY) | Specific (TRACKS 1 & 3) | B | M | PWD | CWD | Color Doppler | Power (Amplitude) Doppler | Other* Combined | Other* Specify |
| Ophthalmic | Ophthalmic | | | | | | | | |
| Fetal Imaging& Other | Fetal | | | | | | | | |
| | Abdominal | | | | | | | | |
| | Intra-operative Specify | | | | | | | | |
| | Intra-operative Neuro | | | | | | | | |
| | Laparoscopic | | | | | | | | |
| | Pediatric | N | N | N | | N | N | Note 1 | Notes 2,4 |
| | Small Organ (specify) | | | | | | | | |
| | Neonatal Cephalic | N | N | N | N | N | N | Note 1 | Notes 2,3,4 |
| | Adult Cephalic | | | | | | | | |
| | Trans-rectal | | | | | | | | |
| | Trans-vaginal | | | | | | | | |
| | Trans-urethral | | | | | | | | |
| | Trans-esoph.(non-Card) | | | | | | | | |
| | Musculo-skeletal (Conventional) | | | | | | | | |
| | Musculo-skeletal (Superficial) | | | | | | | | |
| | Intravascular | | | | | | | | |
| | Other (Ob/GYN) | | | | | | | | |
| Cardiac | Cardiac Adult | | | | | | | | |
| | Cardiac Pediatric | N | N | N | N | N | N | Note 1 | Notes 2,3,4 |
| | Intravascular(Cardiac) | | | | | | | | |
| | Trans-esoph.(Cardiac) | | | | | | | | |
| | Intra-cardiac | | | | | | | | |
| | Other (specify) | | | | | | | | |
| Peripheral Vessel | Peripheral vessel | | | | | | | | |
| | Other (specify) | | | | | | | | |

**N = new indication;** **P = previously cleared by FDA;** **E = added under this appendix**

Note 1: Other Combined includes: B/M; B/PWD; B/THI; M/Color M ; B/Color Doppler; B/Color Doppler/PWD; B/Power Doppler/PWD

Note 2: Tissue Harmonic Imaging. The feature does not use contrast agents

Note 3: TDI Note 4: 3D Note 5: 4D

Note 6: Small Organ: breast, thyroid, testes

Prescription Use ____X___ AND/OR Over-The-Counter Use _____
(Part 21 CFR 801 Subpart D) (21 CFR 807 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of In Vitro Diagnostic Devices (OIVD)

Indications For Use

(Division Sign-Off)
Division of Radiological Devices
Office of In Vitro Diagnostic Device Evaluation and Safety

510K K112602

## Diagnostic Ultrasound Indications for Use Form

Transducer: 6V1 Micro-curved Array
Diagnostic Ultrasound Transducer

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

| Clinical Application | | Mode of Operation | | | | | | | |
|---|---|---|---|---|---|---|---|---|---|
| General (TRACK 1 ONLY) | Specific (TRACKS 1 & 3) | B | M | PWD | CWD | Color Doppler | Power (Amplitude) Doppler | Other* Combined | Other* Specify |
| Ophthalmic | Ophthalmic | | | | | | | | |
| Fetal Imaging& Other | Fetal | | | | | | | | |
| | Abdominal | | | | | | | | |
| | Intra-operative Specify | | | | | | | | |
| | Intra-operative Neuro | | | | | | | | |
| | Laparoscopic | | | | | | | | |
| | Pediatric | | | | | | | | |
| | Small Organ (specify) | | | | | | | | |
| | Neonatal Cephalic | | | | | | | | |
| | Adult Cephalic | | | | | | | | |
| | Trans-rectal | N | N | N | | N | N | Note 1 | Notes 2,4 |
| | Trans-vaginal | N | N | N | | N | N | Note 1 | Notes 2,4 |
| | Trans-urethral | | | | | | | | |
| | Trans-esoph.(non-Card) | | | | | | | | |
| | Musculo-skeletal (Conventional) | | | | | | | | |
| | Musculo-skeletal (Superficial) | | | | | | | | |
| | Intravascular | | | | | | | | |
| | Other (Ob/GYN) | | | | | | | | |
| Cardiac | Cardiac Adult | | | | | | | | |
| | Cardiac Pediatric | | | | | | | | |
| | Intravascular(Cardiac) | | | | | | | | |
| | Trans-esoph.(Cardiac) | | | | | | | | |
| | Intra-cardiac | | | | | | | | |
| | Other (specify) | | | | | | | | |
| Peripheral Vessel | Peripheral vessel | | | | | | | | |
| | Other (specify) | | | | | | | | |

**N = new indication;** **P = previously cleared by FDA;** **E = added under this appendix**

Note 1: Other Combined includes: B/M; B/PWD; B/THI; M/Color M ; B/Color Doppler; B/Color Doppler/PWD; B/Power Doppler/PWD

Note 2: Tissue Harmonic Imaging. The feature does not use contrast agents

Note 3: TDI Note 4: 3D Note 5: 4D

Note 6: Small Organ: breast, thyroid, testes

Prescription Use ___X___ AND/OR Over-The-Counter Use _____
(Part 21 CFR 801 Subpart D) (21 CFR 807 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of In Vitro Diagnostic Devices (OIVD)

Indications For Use

Mary S Pastel
(Division Sign-Off)
Division of Radiological Devices
Office of In Vitro Diagnostic Device Evaluation and Safety

510K K112602

## Diagnostic Ultrasound Indications for Use Form

Transducer: 6V3 Micro-curved Array
Diagnostic Ultrasound Transducer

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

| Clinical Application | | Mode of Operation | | | | | | | |
|---|---|---|---|---|---|---|---|---|---|
| General (TRACK 1 ONLY) | Specific (TRACKS 1 & 3) | B | M | PWD | CWD | Color Doppler | Power (Amplitude) Doppler | Other* Combined | Other* Specify |
| Ophthalmic | Ophthalmic | | | | | | | | |
| Fetal Imaging& Other | Fetal | | | | | | | | |
| | Abdominal | | | | | | | | |
| | Intra-operative Specify | | | | | | | | |
| | Intra-operative Neuro | | | | | | | | |
| | Laparoscopic | | | | | | | | |
| | Pediatric | | | | | | | | |
| | Small Organ (specify) | | | | | | | | |
| | Neonatal Cephalic | | | | | | | | |
| | Adult Cephalic | | | | | | | | |
| | Trans-rectal | N | N | N | | N | N | Note 1 | Notes 2,4 |
| | Trans-vaginal | N | N | N | | N | N | Note 1 | Notes 2,4 |
| | Trans-urethral | | | | | | | | |
| | Trans-esoph.(non-Card) | | | | | | | | |
| | Musculo-skeletal (Conventional) | | | | | | | | |
| | Musculo-skeletal (Superficial) | | | | | | | | |
| | Intravascular | | | | | | | | |
| | Other (Ob/GYN) | | | | | | | | |
| Cardiac | Cardiac Adult | | | | | | | | |
| | Cardiac Pediatric | | | | | | | | |
| | Intravascular(Cardiac) | | | | | | | | |
| | Trans-esoph.(Cardiac) | | | | | | | | |
| | Intra-cardiac | | | | | | | | |
| | Other (specify) | | | | | | | | |
| Peripheral Vessel | Peripheral vessel | | | | | | | | |
| | Other (specify) | | | | | | | | |

**N = new indication;** **P = previously cleared by FDA;** **E = added under this appendix**

Note 1: Other Combined includes: B/M; B/PWD; B/THI; M/Color M ; B/Color Doppler; B/Color Doppler/PWD; B/Power Doppler/PWD

Note 2: Tissue Harmonic Imaging. The feature does not use contrast agents

Note 3: TDI Note 4: 3D Note 5: 4D

Note 6: Small Organ: breast, thyroid, testes

Prescription Use ___X___ AND/OR Over-The-Counter Use _____
(Part 21 CFR 801 Subpart D) (21 CFR 807 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of In Vitro Diagnostic Devices (OIVD)

Indications For Use

Mary S Pastel
(Division Sign-Off)
Division of Radiological Devices
Office of In Vitro Diagnostic Device Evaluation and Safety

510K K112602

## Diagnostic Ultrasound Indications for Use Form

Transducer: EC9-5 Micro-curved Array
Diagnostic Ultrasound Transducer

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

| Clinical Application | | Mode of Operation | | | | | | | |
|---|---|---|---|---|---|---|---|---|---|
| General (TRACK 1 ONLY) | Specific (TRACKS 1 & 3) | B | M | PWD | CWD | Color Doppler | Power (Amplitude) Doppler | Other* Combined | Other* Specify |
| Ophthalmic | Ophthalmic | | | | | | | | |
| Fetal Imaging& Other | Fetal | | | | | | | | |
| | Abdominal | | | | | | | | |
| | Intra-operative Specify | | | | | | | | |
| | Intra-operative Neuro | | | | | | | | |
| | Laparoscopic | | | | | | | | |
| | Pediatric | | | | | | | | |
| | Small Organ (specify) | | | | | | | | |
| | Neonatal Cephalic | | | | | | | | |
| | Adult Cephalic | | | | | | | | |
| | Trans-rectal | N | N | N | | N | N | Note 1 | Notes 2,4 |
| | Trans-vaginal | N | N | N | | N | N | Note 1 | Notes 2,4 |
| | Trans-urethral | | | | | | | | |
| | Trans-esoph.(non-Card) | | | | | | | | |
| | Musculo-skeletal (Conventional) | | | | | | | | |
| | Musculo-skeletal (Superficial) | | | | | | | | |
| | Intravascular | | | | | | | | |
| | Other (Ob/GYN) | | | | | | | | |
| Cardiac | Cardiac Adult | | | | | | | | |
| | Cardiac Pediatric | | | | | | | | |
| | Intravascular(Cardiac) | | | | | | | | |
| | Trans-esoph.(Cardiac) | | | | | | | | |
| | Intra-cardiac | | | | | | | | |
| | Other (specify) | | | | | | | | |
| Peripheral Vessel | Peripheral vessel | | | | | | | | |
| | Other (specify) | | | | | | | | |

**N = new indication; P = previously cleared by FDA; E = added under this appendix**

Note 1: Other Combined includes: B/M; B/PWD; B/THI; M/Color M ; B/Color Doppler; B/Color Doppler/PWD; B/Power Doppler/PWD

Note 2: Tissue Harmonic Imaging. The feature does not use contrast agents

Note 3: TDI Note 4: 3D Note 5: 4D

Note 6: Small Organ: breast, thyroid, testes

Prescription Use ___X___ AND/OR Over-The-Counter Use _____
(Part 21 CFR 801 Subpart D) (21 CFR 807 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of In Vitro Diagnostic Devices (OIVD)

Indications For Use

Mary S Patel
(Division Sign-Off)
Division of Radiological Devices
Office of In Vitro Diagnostic Device Evaluation and Safety

510K K112602

## Diagnostic Ultrasound Indications for Use Form

Transducer: C611 Micro-curved Array
Diagnostic Ultrasound Transducer

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

| Clinical Application | | Mode of Operation | | | | | | | |
|---|---|---|---|---|---|---|---|---|---|
| General (TRACK 1 ONLY) | Specific (TRACKS 1 & 3) | B | M | PWD | CWD | Color Doppler | Power (Amplitude) Doppler | Other* Combined | Other* Specify |
| Ophthalmic | Ophthalmic | | | | | | | | |
| Fetal Imaging& Other | Fetal | | | | | | | | |
| | Abdominal | N | N | N | | N | N | Note 1 | Notes 2,4 |
| | Intra-operative Specify | | | | | | | | |
| | Intra-operative Neuro | | | | | | | | |
| | Laparoscopic | | | | | | | | |
| | Pediatric | N | N | N | | N | N | Note 1 | Notes 2,4 |
| | Small Organ (specify) | | | | | | | | |
| | Neonatal Cephalic | N | N | N | N | N | N | Note 1 | Notes 2,3,4 |
| | Adult Cephalic | | | | | | | | |
| | Trans-rectal | | | | | | | | |
| | Trans-vaginal | | | | | | | | |
| | Trans-urethral | | | | | | | | |
| | Trans-esoph.(non-Card) | | | | | | | | |
| | Musculo-skeletal (Conventional) | | | | | | | | |
| | Musculo-skeletal (Superficial) | | | | | | | | |
| | Intravascular | | | | | | | | |
| | Other (Ob/GYN) | | | | | | | | |
| Cardiac | Cardiac Adult | | | | | | | | |
| | Cardiac Pediatric | N | N | N | N | N | N | Note 1 | Notes 2,3,4 |
| | Intravascular(Cardiac) | | | | | | | | |
| | Trans-esoph.(Cardiac) | | | | | | | | |
| | Intra-cardiac | | | | | | | | |
| | Other (specify) | | | | | | | | |
| Peripheral Vessel | Peripheral vessel | | | | | | | | |
| | Other (specify) | | | | | | | | |

**N = new indication;** **P = previously cleared by FDA;** **E = added under this appendix**

Note 1: Other Combined includes: B/M; B/PWD; B/THI; M/Color M ; B/Color Doppler; B/Color Doppler/PWD; B/Power Doppler/PWD

Note 2: Tissue Harmonic Imaging. The feature does not use contrast agents

Note 3: TDI Note 4: 3D Note 5: 4D

Note 6: Small Organ: breast, thyroid, testes

Prescription Use ___X___ AND/OR Over-The-Counter Use _____
(Part 21 CFR 801 Subpart D) (21 CFR 807 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of In Vitro Diagnostic Devices (OIVD)

Indications For Use

(Division Sign-Off)
Division of Radiological Devices
Office of In Vitro Diagnostic Device Evaluation and Safety

510K K112602

## Diagnostic Ultrasound Indications for Use Form

Transducer: C362 Curved Array
Diagnostic Ultrasound Transducer

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

| Clinical Application | | Mode of Operation | | | | | | | |
|---|---|---|---|---|---|---|---|---|---|
| General (TRACK 1 ONLY) | Specific (TRACKS 1 & 3) | B | M | PWD | CWD | Color Doppler | Power (Amplitude) Doppler | Other* Combined | Other* Specify |
| Ophthalmic | Ophthalmic | | | | | | | | |
| Fetal Imaging & Other | Fetal | N | N | N | | N | N | Note 1 | Notes 2,4 |
| | Abdominal | N | N | N | | N | N | Note1 | Notes 2,4 |
| | Intra-operative Specify | | | | | | | | |
| | Intra-operative Neuro | | | | | | | | |
| | Laparoscopic | | | | | | | | |
| | Pediatric | | | | | | | | |
| | Small Organ (specify) | | | | | | | | |
| | Neonatal Cephalic | | | | | | | | |
| | Adult Cephalic | | | | | | | | |
| | Trans-rectal | | | | | | | | |
| | Trans-vaginal | | | | | | | | |
| | Trans-urethral | | | | | | | | |
| | Trans-esoph.(non-Card) | | | | | | | | |
| | Musculo-skeletal (Conventional) | | | | | | | | |
| | Musculo-skeletal (Superficial) | | | | | | | | |
| | Intravascular | | | | | | | | |
| | Other (Ob/GYN) | N | N | N | | N | N | Note1 | Notes 2,4 |
| Cardiac | Cardiac Adult | | | | | | | | |
| | Cardiac Pediatric | | | | | | | | |
| | Intravascular(Cardiac) | | | | | | | | |
| | Trans-esoph.(Cardiac) | | | | | | | | |
| | Intra-cardiac | | | | | | | | |
| | Other (specify) | | | | | | | | |
| Peripheral Vessel | Peripheral vessel | | | | | | | | |
| | Other (specify) | | | | | | | | |

**N = new indication;** **P = previously cleared by FDA;** **E = added under this appendix**

Note 1: Other Combined includes: B/M; B/PWD; B/THI; M/Color M ; B/Color Doppler; B/Color Doppler/PWD; B/Power Doppler/PWD

Note 2: Tissue Harmonic Imaging. The feature does not use contrast agents

Note 3: TDI Note 4: 3D Note 5: 4D

Note 6: Small Organ: breast, thyroid, testes

Prescription Use ____X___ AND/OR Over-The-Counter Use _____
(Part 21 CFR 801 Subpart D) (21 CFR 807 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of In Vitro Diagnostic Devices (OIVD)

Indications For Use

Mary S Pastel
(Division Sign-Off)
Division of Radiological Devices
Office of In Vitro Diagnostic Device Evaluation and Safety

510K K112602

## Diagnostic Ultrasound Indications for Use Form

Transducer: C344 Curved Array
Diagnostic Ultrasound Transducer

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

| Clinical Application | | Mode of Operation | | | | | | | |
|---|---|---|---|---|---|---|---|---|---|
| General (TRACK 1 ONLY) | Specific (TRACKS 1 & 3) | B | M | PWD | CWD | Color Doppler | Power (Amplitude) Doppler | Other* Combined | Other* Specify |
| Ophthalmic | Ophthalmic | | | | | | | | |
| Fetal Imaging& Other | Fetal | N | N | N | | N | N | Note 1 | Notes 2, 4 |
| | Abdominal | N | N | N | | N | N | Note 1 | Notes 2, 4 |
| | Intra-operative Specify | | | | | | | | |
| | Intra-operative Neuro | | | | | | | | |
| | Laparoscopic | | | | | | | | |
| | Pediatric | | | | | | | | |
| | Small Organ (specify) | | | | | | | | |
| | Neonatal Cephalic | | | | | | | | |
| | Adult Cephalic | | | | | | | | |
| | Trans-rectal | | | | | | | | |
| | Trans-vaginal | | | | | | | | |
| | Trans-urethral | | | | | | | | |
| | Trans-esoph.(non-Card) | | | | | | | | |
| | Musculo-skeletal (Conventional) | | | | | | | | |
| | Musculo-skeletal (Superficial) | | | | | | | | |
| | Intravascular | | | | | | | | |
| | Other (Ob/GYN) | N | N | N | | N | N | Note 1 | Notes 2, 4 |
| Cardiac | Cardiac Adult | | | | | | | | |
| | Cardiac Pediatric | | | | | | | | |
| | Intravascular(Cardiac) | | | | | | | | |
| | Trans-esoph.(Cardiac) | | | | | | | | |
| | Intra-cardiac | | | | | | | | |
| | Other (specify) | | | | | | | | |
| Peripheral Vessel | Peripheral vessel | | | | | | | | |
| | Other (specify) | | | | | | | | |

**N = new indication; P = previously cleared by FDA; E = added under this appendix**

Note 1: Other Combined includes: B/M; B/PWD; B/THI; M/Color M ; B/Color Doppler; B/Color Doppler/PWD; B/Power Doppler/PWD

Note 2: Tissue Harmonic Imaging. The feature does not use contrast agents

Note 3: TDI Note 4: 3D Note 5: 4D

Note 6: Small Organ: breast, thyroid, testes

Prescription Use ___X___ AND/OR Over-The-Counter Use _____
(Part 21 CFR 801 Subpart D) (21 CFR 807 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of In Vitro Diagnostic Devices (OIVD)

Indications For Use

(Division Sign-Off)
Division of Radiological Devices
Office of In Vitro Diagnostic Device Evaluation and Safety

510K K112602

## Diagnostic Ultrasound Indications for Use Form

Transducer: VC6-2 Curved Array
Diagnostic Ultrasound Transducer

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

| Clinical Application | | Mode of Operation | | | | | | | |
|---|---|---|---|---|---|---|---|---|---|
| General (TRACK 1 ONLY) | Specific (TRACKS 1 & 3) | B | M | PWD | CWD | Color Doppler | Power (Amplitude) Doppler | Other* Combined | Other* Specify |
| Ophthalmic | Ophthalmic | | | | | | | | |
| Fetal Imaging & Other | Fetal | N | N | N | | N | N | Note 1 | Notes 2,4,5 |
| | Abdominal | N | N | N | | N | N | Note 1 | Notes 2,4,5 |
| | Intra-operative Specify | | | | | | | | |
| | Intra-operative Neuro | | | | | | | | |
| | Laparoscopic | | | | | | | | |
| | Pediatric | | | | | | | | |
| | Small Organ (specify) | | | | | | | | |
| | Neonatal Cephalic | | | | | | | | |
| | Adult Cephalic | | | | | | | | |
| | Trans-rectal | | | | | | | | |
| | Trans-vaginal | | | | | | | | |
| | Trans-urethral | | | | | | | | |
| | Trans-esoph.(non-Card) | | | | | | | | |
| | Musculo-skeletal (Conventional) | | | | | | | | |
| | Musculo-skeletal (Superficial) | | | | | | | | |
| | Intravascular | | | | | | | | |
| | Other (Ob/GYN) | N | N | N | | N | N | Note 1 | Notes 2,4,5 |
| Cardiac | Cardiac Adult | | | | | | | | |
| | Cardiac Pediatric | | | | | | | | |
| | Intravascular(Cardiac) | | | | | | | | |
| | Trans-esoph.(Cardiac) | | | | | | | | |
| | Intra-cardiac | | | | | | | | |
| | Other (specify) | | | | | | | | |
| Peripheral Vessel | Peripheral vessel | | | | | | | | |
| | Other (specify) | | | | | | | | |

**N = new indication; P = previously cleared by FDA; E = added under this appendix**

Note 1: Other Combined includes: B/M; B/PWD; B/THI; M/Color M ; B/Color Doppler; B/Color Doppler/PWD; B/Power Doppler/PWD

Note 2: Tissue Harmonic Imaging. The feature does not use contrast agents

Note 3: TDI Note 4: 3D Note 5: 4D

Note 6: Small Organ: breast, thyroid, testes

Prescription Use ___X___ AND/OR Over-The-Counter Use _____
(Part 21 CFR 801 Subpart D) (21 CFR 807 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of In Vitro Diagnostic Devices (OIVD)

Mary S Patel
(Division Sign-Off)
Division of Radiological Devices
Office of In Vitro Diagnostic Device Evaluation and Safety

510K K112602

## Diagnostic Ultrasound Indications for Use Form

Transducer: L743 Linear Array
Diagnostic Ultrasound Transducer

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

| Clinical Application | | Mode of Operation | | | | | | | |
|---|---|---|---|---|---|---|---|---|---|
| General (TRACK 1 ONLY) | Specific (TRACKS 1 & 3) | B | M | PWD | CWD | Color Doppler | Power (Amplitude) Doppler | Other* Combined | Other* Specify |
| Ophthalmic | Ophthalmic | | | | | | | | |
| Fetal Imaging& Other | Fetal | | | | | | | | |
| | Abdominal | | | | | | | | |
| | Intra-operative Specify | | | | | | | | |
| | Intra-operative Neuro | | | | | | | | |
| | Laparoscopic | | | | | | | | |
| | Pediatric | | | | | | | | |
| | Small Organ (specify) | N | N | N | | N | N | Note 1 | Notes 2, 4 |
| | Neonatal Cephalic | | | | | | | | |
| | Adult Cephalic | | | | | | | | |
| | Trans-rectal | | | | | | | | |
| | Trans-vaginal | | | | | | | | |
| | Trans-urethral | | | | | | | | |
| | Trans-esoph.(non-Card) | | | | | | | | |
| | Musculo-skeletal (Conventional) | N | N | N | | N | N | Note 1 | Notes 2, 4 |
| | Musculo-skeletal (Superficial) | N | N | N | | N | N | Note 1 | Notes 2, 4 |
| | Intravascular | | | | | | | | |
| | Other (Ob/GYN) | | | | | | | | |
| Cardiac | Cardiac Adult | | | | | | | | |
| | Cardiac Pediatric | | | | | | | | |
| | Intravascular(Cardiac) | | | | | | | | |
| | Trans-esoph.(Cardiac) | | | | | | | | |
| | Intra-cardiac | | | | | | | | |
| | Other (specify) | | | | | | | | |
| Peripheral Vessel | Peripheral vessel | N | N | N | | N | N | Note 1 | Notes 2, 4 |
| | Other (specify) | | | | | | | | |

**N = new indication;** **P = previously cleared by FDA;** **E = added under this appendix**

Note 1: Other Combined includes: B/M; B/PWD; B/THI; M/Color M ; B/Color Doppler; B/Color Doppler/PWD; B/Power Doppler/PWD

Note 2: Tissue Harmonic Imaging. The feature does not use contrast agents

Note 3: TDI Note 4: 3D Note 5: 4D

Note 6: Small Organ: breast, thyroid, testes

Prescription Use ___X___ AND/OR Over-The-Counter Use _____
(Part 21 CFR 801 Subpart D) (21 CFR 807 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of In Vitro Diagnostic Devices (OIVD)

Indications For Use

Mary S Patel
(Division Sign-Off)
Division of Radiological Devices
Office of In Vitro Diagnostic Device Evaluation and Safety

510K K112602

## Diagnostic Ultrasound Indications for Use Form

Transducer: L741 Linear Array
Diagnostic Ultrasound Transducer

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

| Clinical Application | | Mode of Operation | | | | | | | |
|---|---|---|---|---|---|---|---|---|---|
| General (TRACK 1 ONLY) | Specific (TRACKS 1 & 3) | B | M | PWD | CWD | Color Doppler | Power (Amplitude) Doppler | Other* Combined | Other* Specify |
| Ophthalmic | Ophthalmic | | | | | | | | |
| Fetal Imaging& Other | Fetal | | | | | | | | |
| | Abdominal | | | | | | | | |
| | Intra-operative Specify | | | | | | | | |
| | Intra-operative Neuro | | | | | | | | |
| | Laparoscopic | | | | | | | | |
| | Pediatric | | | | | | | | |
| | Small Organ (specify) | N | N | N | | N | N | Note 1 | Notes 2, 4 |
| | Neonatal Cephalic | | | | | | | | |
| | Adult Cephalic | | | | | | | | |
| | Trans-rectal | | | | | | | | |
| | Trans-vaginal | | | | | | | | |
| | Trans-urethral | | | | | | | | |
| | Trans-esoph.(non-Card) | | | | | | | | |
| | Musculo-skeletal (Conventional) | N | N | N | | N | N | Note 1 | Notes 2, 4 |
| | Musculo-skeletal (Superficial) | | | | | | | | |
| | Intravascular | | | | | | | | |
| | Other (Ob/GYN) | | | | | | | | |
| Cardiac | Cardiac Adult | | | | | | | | |
| | Cardiac Pediatric | | | | | | | | |
| | Intravascular(Cardiac) | | | | | | | | |
| | Trans-esoph.(Cardiac) | | | | | | | | |
| | Intra-cardiac | | | | | | | | |
| | Other (specify) | | | | | | | | |
| Peripheral Vessel | Peripheral vessel | N | N | N | | N | N | Note 1 | Notes 2, 4 |
| | Other (specify) | | | | | | | | |

N = new indication; P = previously cleared by FDA; E = added under this appendix

Note 1: Other Combined includes: B/M; B/PWD; B/THI; M/Color M ; B/Color Doppler; B/Color Doppler/PWD; B/Power Doppler/PWD

Note 2: Tissue Harmonic Imaging. The feature does not use contrast agents

Note 3: TDI Note 4: 3D Note 5: 4D

Note 6: Small Organ: breast, thyroid, testes

Prescription Use ___X___ AND/OR Over-The-Counter Use _____
(Part 21 CFR 801 Subpart D) (21 CFR 807 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of In Vitro Diagnostic Devices (OIVD)

Indications For Use

Mary S Pastel
(Division Sign-Off)
Division of Radiological Devices
Office of In Vitro Diagnostic Device Evaluation and Safety

510K K112602

## Diagnostic Ultrasound Indications for Use Form

Transducer: L742 Linear Array
Diagnostic Ultrasound Transducer

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

| Clinical Application | | Mode of Operation | | | | | | | |
|---|---|---|---|---|---|---|---|---|---|
| General (TRACK 1 ONLY) | Specific (TRACKS 1 & 3) | B | M | PWD | CWD | Color Doppler | Power (Amplitude) Doppler | Other* Combined | Other* Specify |
| Ophthalmic | Ophthalmic | | | | | | | | |
| Fetal Imaging& Other | Fetal | | | | | | | | |
| | Abdominal | | | | | | | | |
| | Intra-operative Specify | | | | | | | | |
| | Intra-operative Neuro | | | | | | | | |
| | Laparoscopic | | | | | | | | |
| | Pediatric | | | | | | | | |
| | Small Organ (specify) | N | N | N | | N | N | Note 1 | Notes 2, 4 |
| | Neonatal Cephalic | | | | | | | | |
| | Adult Cephalic | | | | | | | | |
| | Trans-rectal | | | | | | | | |
| | Trans-vaginal | | | | | | | | |
| | Trans-urethral | | | | | | | | |
| | Trans-esoph.(non-Card) | | | | | | | | |
| | Musculo-skeletal (Conventional) | N | N | N | | N | N | Note 1 | Notes 2, 4 |
| | Musculo-skeletal (Superficial) | N | N | N | | N | N | Note 1 | Notes 2, 4 |
| | Intravascular | | | | | | | | |
| | Other (Ob/GYN) | | | | | | | | |
| Cardiac | Cardiac Adult | | | | | | | | |
| | Cardiac Pediatric | | | | | | | | |
| | Intravascular(Cardiac) | | | | | | | | |
| | Trans-esoph.(Cardiac) | | | | | | | | |
| | Intra-cardiac | | | | | | | | |
| | Other (specify) | | | | | | | | |
| Peripheral Vessel | Peripheral vessel | N | N | N | | N | N | Note 1 | Notes 2, 4 |
| | Other (specify) | | | | | | | | |

**N = new indication; P = previously cleared by FDA; E = added under this appendix**

Note 1: Other Combined includes: B/M; B/PWD; B/THI; M/Color M ; B/Color Doppler; B/Color Doppler/PWD; B/Power Doppler/PWD

Note 2: Tissue Harmonic Imaging. The feature does not use contrast agents

Note 3: TDI Note 4: 3D Note 5: 4D

Note 6: Small Organ: breast, thyroid, testes

Prescription Use ___X___ AND/OR Over-The-Counter Use _____
(Part 21 CFR 801 Subpart D) (21 CFR 807 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of In Vitro Diagnostic Devices (OIVD)

Indications For Use

Mary S Pastel
(Division Sign-Off)
Division of Radiological Devices
Office of In Vitro Diagnostic Device Evaluation and Safety

510K K112602